




## BD MAX™ ExK™ DNA-2

(CSF or Dry Swabs)

**REF** 442819

**IVD**



For use with the BD MAX™ System

### INTENDED USE

The BD MAX™ ExK™ DNA-2 is intended to extract DNA from cerebral spinal fluid (CSF) or dry swab specimens. The DNA Extraction Tube contains a Specimen Processing Control (SPC) sequence from *Drosophila melanogaster* (GenBank: AC246436; nucleotides 35779 to 35978) cloned in a pUC119 derivative (GenBank: U07650.1). The BD MAX™ ExK™ DNA-2 has not been validated for use with any specific analytical test method.

### PRINCIPLES OF THE PROCEDURE

The specimen is mixed with the BD MAX™ Sample Buffer Tube and processed using the BD MAX™ System. The BD MAX™ System automates DNA extraction and concentration. No operator intervention is necessary once the specimen is loaded onto the BD MAX™ System. The BD MAX™ System contains a combination of lytic and extraction reagents designed to perform cell lysis, DNA extraction and removal of inhibitors. Following cell lysis, the released DNA is captured by magnetic affinity beads. The beads, with the bound DNA, are washed and the DNA is eluted using 12.5 µL of Elution Buffer. The solution may be prepared for PCR using neutralization buffer from the DNA Strip. The eluted DNA may be used for applications on the BD MAX™ System or another PCR Amplification/Detection System.

### REAGENTS AND SUPPLIES

REF	COMPONENT NAME	QUANTITY
442819	DNA Extraction Tube DNA Sample Buffer Tube SB-2 DNA Strip Septum Conical Tube	24

### EQUIPMENT AND MATERIALS REQUIRED BUT NOT PROVIDED

- BD MAX™ System: REF 441916 or 441927
- Micropipettes (accurate between 2 and 1000 µL)
- Aerosol resistant micropipette tips
- Disposable gloves/lab coat
- Specimen Collection Container(s)
- Dry Swabs

### WARNINGS AND PRECAUTIONS

- This kit is for *in vitro* diagnostic use.
- Do not use this kit if the outer carton safety seal is broken.
- Do not use this kit if the packaging is damaged upon arrival.
- Do not use reagents if their protective pouches are open or broken upon arrival.
- Do not use reagents if desiccant is not present or broken inside protective pouches.
- Do not remove desiccant from protective pouches.
- Close protective pouches promptly with the zip seal after each use. Remove any excess air in the pouches prior to sealing.
- Do not use reagents if their foil seal has been opened or damaged (DNA Extraction Tube, DNA Strip).

- Proceed with caution when using chemical solutions as Master Mix and Extraction Tube barcode readability may be altered.
- Check DNA Strips for proper liquid fills (ensure that the liquids are at the bottom of all three (3) buffer tubes; if necessary gently tap the DNA Strips until the liquids move to the bottom of tubes, see Figure 1).
- Check DNA Strips to ensure that all four (4) pipette tips are present (see Figure 1).
- Protect reagents against heat and humidity. Prolonged exposure to humidity may affect product performance.
- Reagents are not interchangeable between lots.
- Never pool reagents from different tubes even if they are from the same lot.
- Do not interchange caps among reagents as contamination may occur and compromise test results.
- Do not use reagents after expiration date.
- Avoid microbial and deoxyribonuclease (DNase) contamination of reagents at all times. The use of sterile DNase-free disposable aerosol resistant or positive displacement pipette tips is recommended. Use a new tip for each specimen.
- Always handle specimens as if they are infectious and in accordance with safe laboratory procedures such as those described in *Biosafety in Microbiological and Biomedical Laboratories*<sup>1</sup> and in CLSI Document M29<sup>2</sup>.
- Wear personal protective equipment and non-powdered disposable gloves while handling all reagents. Wash hands thoroughly after performing the test.
- Do not pipette by mouth, smoke, drink, or eat in areas where specimens or kit reagents are being handled.
- Dispose of unused reagents and waste in accordance with country, federal, provincial, state and local regulations.
- Consult the BD MAX™ System User's Manual (REF 441939 or 441942) for additional warnings, precautions and procedures.

### STORAGE AND STABILITY

- BD MAX™ ExK™ DNA-2 is stable at 2-25°C through the stated expiration date. The expiration date applies only to unopened kits, stored as directed. Do not use expired components.
- BD MAX™ ExK™ DNA-2 Extraction Tubes are provided in sealed pouches. To protect product from humidity, immediately re-seal after opening. Reagent tubes are stable for up to 7 days at 2-8 °C after initial opening and re-sealing.

### INSTRUCTIONS FOR USE

#### Specimen Collection

- Collect specimen and label appropriately.
- Collected specimens should be maintained per the conditions validated by the user.
- Proceed to Extraction Preparation.

#### Extraction Preparation

**Note:** The following extraction preparation procedures are provided as guidance. Some specimens may require pre-processing. Application-specific extraction preparation procedures should be developed and validated by the user.

- CSF:** Pipette 200 µL of specimen into a BD MAX™ DNA Sample Buffer Tube SB-2 and close the tube with a septum cap. Ensure complete mixing by vortexing the sample.
- Dry Swab:** Insert swab into a BD MAX™ DNA Sample Buffer Tube SB-2 and swirl at least three times. Lift the swab out of the liquid and press against the side of the Sample Buffer Tube to expel residual specimen. Completely remove the swab from the tube and discard appropriately. Close the Sample Buffer Tube with a septum cap and ensure complete mixing by vortexing the sample.
- Proceed to BD MAX™ Operation.

#### BD MAX™ Operation

Refer to BD MAX™ System User's Manual (REF 441939 or 441942)

for programming and setup instructions.

1. For each specimen to be tested, place one (1) DNA Strip on the BD MAX System Rack.
2. Assemble BD MAX™ DNA Strips.
  - a. Snap the appropriate DNA Extraction Tube into Position 1 of the DNA Strip, as shown in Figure 1. The tube is fully seated in the strip when a 'click' is heard. Only one (1) DNA Extraction Tube should be used with each DNA Strip.
  - b. If performing DNA extraction only, snap the Conical Tube into Position 3 of the DNA Strip, as shown in Figure 1. The tube is fully seated in the strip when a 'click' is heard. Proceed to Step 3 below.
  - c. If the BD MAX™ System is also being used for DNA amplification using BD MAX™ DNA MMK (SPC) or BD MAX™ DNA MMK (REF 442829 or 442848), follow the instructions for use accompanying that product. Snap the MM tube into Position 2 and snap the Conical tube into Position 3 of the DNA Strip, as shown in Figure 1. The tubes are fully seated in the strip when a 'click' is heard. Pipette 12.5 µL of primer and probe mixture (at a final concentration of 2X) into the Conical Tube in Position 3 of the DNA Strip. Proceed to Step 3 below.
  - d. If the BD MAX™ System is being used for DNA amplification with a user supplied master mix, snap a Conical Tube into Position 3 of the DNA Strip, as shown in Figure 1. The tube is fully seated in the strip when a 'click' is heard. Prepare liquid primer and probe mixture for target(s) and user-provided sample processing control. Pipette a total of 12.5 µL of the user supplied master mix, including primers and probes (at a final concentration of 2X) into the Conical Tube in Position 3 of the DNA Strip. Add user-provided sample processing control to the Sample Buffer Tube. Proceed to Step 3 below.
3. Select the 'Work List' tab in the 'Run' screen on the BD MAX™ System monitor.
4. Enter the specimen/patient identification number into the BD MAX™ System, using either the barcode scanner or manual entry. Start with Position 1 of Rack A. (Rack A is positioned on the left side of the BD MAX™ System and Rack B is on the right, Figure 2).
5. Enter the Sample Buffer Tube barcode corresponding to each specimen/patient using the barcode scanner or manual entry. Start with Position 1 of Rack A and ensure that the specimen/patient ID and the Sample Buffer Tube are accurately matched.
6. Place the Sample Buffer Tube (containing the specimen) on the BD MAX™ System Rack. Start with Position 1 of Rack A.
7. Repeat Steps (1-6) for all specimens. Always move from Position 1 through Position 12 on each Rack, ensuring that no positions are skipped.
8. If the BD MAX™ System is also being used for DNA amplification, include additional disposables, as required.
9. Load Rack(s) into the BD MAX™ System (Figure 2). Ensure that the placement of the Rack(s) corresponds to the 'Work List' definition.
10. Close the BD MAX™ System door and press Start to begin the test run.

Figure 1. DNA Strip





Figure 2. BD MAX™ System



Rack A

Rack B

#### LIMITATIONS OF THE PROCEDURE

1. The BD MAX™ ExK™ DNA-2 can only be used on the BD MAX™ System by trained personnel.
2. Use of the BD MAX™ ExK™ DNA-2 for clinical specimen types other than those specified have not been validated.
3. The user must validate the selected application of this product according to country, federal, provincial, state, local and/or accrediting organization guidelines, regulations and standards.

#### REFERENCES

- 1 Centers for Disease Control and Prevention. *Biosafety in Microbiological and Biomedical Laboratories*. Chosewood LC and Wison DE (eds) (2009). HHS Publication number (CDC) 21-1112.
- 2 Clinical and Laboratory Standards Institute. Protection of laboratory workers from occupationally acquired infections; Approved Guideline – Document M29 (Refer to the latest edition).

SYMBOL	MEANING	SYMBOL	MEANING
	Catalog number		Batch code
	Use by		Temperature limitation
	Contains sufficient for "n" tests		Consult instructions for use
	Reseal pouch after use		Keep dry
	Manufacturer		In vitro Diagnostic Use



Technical Information: In the United States, telephone Technical Services, toll-free (800) 638-8663.

GeneOhm Sciences Canada, Inc., 2555 boul. du Parc-Technologique, Québec, Qc, Canada G1P 4S5

BD, BD Logo and all other trademarks are property of Becton, Dickinson and Company.

© 2012 BD. Made in Canada.